Inspection Certificate
Abnahmeprüfungszeugnis DIN EN 10204.3.1
Certificat De reception
Certificado di collaudo
Keuringsrapport

Batch Number	**2609035**	**Product Name**	Zyglo ZL-56 Water Washable Fluorescent Penetrant
Date of Manufacture	28/09/2026	**B.B.E.**	09/2031

Specification: Specification

Specification
We hereby certify that when tested at the time of manufacture, the above material:
1. Meets the requirements of and has been tested for sulfur and halogens according to:
a) ASME Boiler and Pressure Vessel Code, 2025 Edition, Section V, Non-destructive Examination.

b) ASTM E-165/E-165M-23, Paragraph 7.1.

c) MIL-STD-2132E, March 29, 2016, Paragraph 6.1.3.

2. Meets the requirements of Rolls Royce RRP 58003 (CSS 232), AMS 2644J and ASTM E 1417/E 1417M-21 EDT1 Paragraph 5.1& 6.5.1.

Penetrant brightness tested according to ASTM E 1135.

3. We further certify that the material does not contain mercury as a basic element and no mercury bearing equipment was used in its manufacture.

Specification: Ion Testing Results
Test results obtained were as follows:

Test	Section	Limit	Result
Sulphur Content (S)			0.0054%
Halogen Content (F + Cl)			0.0299%

Specification: AMS 2644J Penetrant Test Results
Type 1 Level 4 Penetrant

When sampled and tested according to paragraph 4.3.2 section 4.2.2.1 of AMS 2644J the following results were obtained:

Test	Section	Limit	Result
Flashpoint	3.3.3	Complies with standard	135°C
Viscosity	3.3.4	Complies with standard	18.14 mm2/s (cSt)
Fluorescent Brightness	3.3.8.3.3	Complies with standard	122%
Water Tolerance	3.3.8.5	Complies with standard	27.65%
Penetrant Removability	3.3.8.7	Complies with standard	Passed
Water Content	3.3.8.8	Complies with standard	0.10%

--- EOR ---

Inspection Certificate

Abnahmeprüfungszeugnis DIN EN 10204.3.1
Certificat De reception
Certificado di collaudo
Keuringsrapport

Prepared by



Approved by



Notes:

1. Our batch number appears on the label of bulk containers. Aerosols have batch numbers printed on bottom of the container.
2. Most specifications require test results stated in percent, but some require parts per million (ppm). To convert "percent" figures to "parts per million" move the decimal four places to the right.
3. MIL-STD-271, MIL-STD-2132 and ASME Sec V, all require that materials be subject to a procedure to evaporate off volatile solvents before analysis for Sulfur and Halogens. According to these specifications, only those residues higher than 0.005 g/100ml shall be analysed for Sulfur and Halogens. Lower residues shall be reported.
4. The above certification gives the results obtained at the time of manufacture. Age and use may alter the properties of any material.

Magnaflux (A Division of ITW Ltd),
Faraday Road, South Dorcan Industrial Estate,
Swindon, Wiltshire, SN3 5HE, UK
Tel: +44 (0)1793 524566
Fax: +44 (0)1793 490459
Email: sales.eu@magnaflux.com
www.magnaflux.eu